



1010 WEST CHESTNUT, PO BOX 947 CHANUTE, KS 66720-0947

CERTIFICATE OF ACCURACY

This is to certify that on the 30 day of March 2015 the instrument below was tested and meets the manufacturer's specifications.

Manufacturer: KUSTOM SIGNALS, INC.

Model Falcon HR

Serial Number:

Transmitter Frequency Measured

Indicator SH 09600 24.129 GHz

(Frequency tolerance allowed: K-Band 24.025 - 24.225 GHz, Ka-Band 35.400 - 35.600 GHz)

MPH KM/H _____ KNOTS _____

Calibration Speed	Speed Registered
35	35
50	50
65	65

HEWLETT PACKARD FREQUENCY COUNTER
 FREQUENCIES MODEL NO. 5352B
 CHECKED USING: SERIAL NO. 2816A00487

The instrument(s) used to certify the accuracy of the above radar has been calibrated within the previous year and is traceable to the National Institute of Standards and Technology.


Calibration performed by Technician Tammy Radtch

Calibration verified by: Kevin J. Unrein

FCC license number: PG-17-21280

Expiration date: NO EXPIRATION DATE

Subscribed and sworn to me this 30 day of March 2015



Lori A. Shue
NOTARY PUBLIC